

SoftPro™ In Bed AFO

Fitting and Care instructions



Applying the Brace

1. If necessary, heat mold the semi-rigid insert if the plantar flexion or inversion /eversion is greater than 15° from neutral. Mold to within 10° of comfortable end range stretch for plantar flexion, and within 5° of inversion/eversion end range and to accommodate any rotation of the foot to significantly reduce any possibility of pressure on the ankle/foot with device use.
2. Open the closures so that the device is ready to apply after passive stretching of the ankle/foot.
3. With the patient in a supine position, passively stretch the ankle/foot. Flex the knee if necessary to 45° or as close to 45° as comfort will allow obtaining optimal positioning of the ankle/foot post stretch. Avoid a stretch reflex while passively stretching the joint. Passive stretch must address the plantar flexion, inversion /eversion, and rotation of the foot. Slowly and gently use sub-maximal passive stretching to point of noticeable resistance only (no discomfort). Hold for a minute to allow the extension release of the affected joint(s).
4. Place the patient's foot in the AFO ensuring that the heel of the foot is at the apex of the heel of the AFO.
5. Holding the AFO in place, bring the loop side of the foot portion the liner over the top of the foot. Bring the opposite side of the foot liner over the top of the foot and secure the hook and loop closure. A finger should comfortably slip under both edges of the closure to ensure that the closure is not secured too tightly.
6. Bring the loop side of the calf portion of the liner over the calf. Bring the opposite side of the liner over the calf and secure the hook and loop closure. A finger should comfortably slip under both edges of the closure to ensure that the closure is not secured too tightly.
7. Bring the unsecured end of the anterior ankle strap over the top of the ankle aligning the strap with the black loop landing zone on the top of the foot. Continue taking the anterior ankle strap over the entire foot, loop the strap through the open slot at the bottom of the heel, and secure the strap back onto itself. The strap should be snug, but not tight. If needed, adjust both ends of the anterior strap if the strap needs to be centered to maintain a good closure.
8. Bring the hook closure at the top of foot liner portion of the liner over the anterior strap and attach the closure over the strap.
9. Check to insure that the heel is floating free in the brace with clearance off of the plastic heel of the brace.
10. Move the hip control bar if needed to the left or right side to control or prevent internal or external rotation of the hip when the resident is in bed

11. Attach the removable elastic pull strap by placing the center of the strap under the top of the toe plate and securing the strap centered under the sole of the foot in the boot. Take each end of the strap up through the D ring on the top back of the AFO and secure the strap onto itself at the desired therapeutic tension. Tighten both straps to treat plantar flexion. Tighten only the lateral strap to treat inversion of the ankle/foot. Tighten only the medial strap to treat eversion of the ankle/foot.
12. If using the elastic pull strap, tension should initiate a GENTLE stretch of the ankle/foot in the corrective direction. There should be no pain or discomfort. The patient should experience a gentle stretching for several minutes. The stretching sensation should diminish and become more comfortable after several minutes of wear. White out or an indelible marker can be used to mark the pull straps on each side to delineate pull strap settings for caregivers.
13. Determine wearing schedule per therapy evaluation and physician's order.
14. Use the device in the recumbent position or for assisted weight bearing and initial gait training.
15. Prior to application, inspect the device. If there are no problems with the device, inspect the ankle/foot. Do not apply the device if there is significant redness on the skin that would come in contact with the device.
16. Upon device removal, closely inspect for skin integrity. Notify the appropriate staff member(s) immediately and document any significant redness or signs of adverse pressure or shear. Discontinue device use until the skin integrity issues are resolved, and the orthosis is adjusted, or the wearing schedule is modified to eliminate potential skin integrity problems.
17. Always follow medical device protocols if the patient is transferred to acute care or another healthcare facility. Send the related orthotic care plan and any accessories along with the orthotic device(s) with the patient if the orthotic will be going with the patient.
18. A wearing time adjustment is necessary every time there is a significant disruption in wearing schedule. Device tolerance must be reintroduced gradually and noted in the patient's care plan.
19. Check device settings for continued application of the desired amount of dorsi-flexion stretch at least once a month or whenever the patient is not experiencing a gentle stretch sensation post application. Re-adjust the orthosis to maintain 5 to 10° of dorsi-flexion beyond the point of resistance if treating plantar flexion contractures.

LAUNDRY INSTRUCTIONS

1. Always remove soft cover from frame before washing.
2. Close all hook and loop attachments on soft cover and place in enclosed laundry bag.
3. Hand or machine wash, gentle cycle with mild detergent. **DO NOT USE COMMERCIAL WASHERS OR HOT WATER.**
4. No bleach or fabric softener
5. Air dry.